

Manual Supplement

Manual Title:	110,111,112 Calibration Info.	Supplement Issue:	1
Part Number:	1608592	Issue Date:	8/01
Print Date:	March 2001	Page Count:	1
Revision/Date:			

This supplement contains information necessary to ensure the accuracy of the above manual. Enter the corrections in the manual if either one of the following conditions exist:

1. The revision letter stamped on the indicated PCA is equal to or higher than that given with each change.
2. No revision letter is indicated at the beginning of the change.

Change #1

On page 12, Table 2, replace the \bar{V} display limits with:

Input Value	Frequency (sine wave)	Display Lower Limits	Display Upper Limits
0 V	0 Hz	-0.002	0.002
5 V		4.963	5.037
-50 V		-49.63	-50.37
600 V		595.6	604.4