

Optical Wavelength Laboratories

Customer Name John Smith
 Report Number 2917
 Meter Serial Number FO-10396
 Source Serial Number LO-10447
 Transfer Standard ID LO-11001

Date of Calibration February 14, 2001
 Date of Next Calibration February 14, 2002
 Calibrated by Joe Testerman
 Temp. (°F) 72.1 Rel. Hum. (%) 30%
 Technician Signature John Hancock

NIST # **814342**

Optical Light Source Certificate of Calibration

Wavelength	850nm	1300/1310nm	1550nm
Coupled Power	-10.00 dBm	-10.00 dBm	-10.00 dBm
Fiber Core Diameter	N/A μM	8 μM	9 μM
Connector Style	FC	FC	FC

Optical Attenuator Certificate of Calibration

Wavelength	850nm	1300/1310nm	1550nm
Transfer Standard	N/A dBm	-10.00 dBm	-10.00 dBm
Uncertainty	N/A	N/A	N/A
Calibration	N/A	-10.00 dBm	-10.00 dBm

Range Check	-10dBm -10.00	-20dBm -20.00	-30dBm -30.00
1310 nm	-40dBm -40.00	-50dBm -50.00	-60dBm N/A

N.I.S.T. TRACEABLE